

Empore™

Technetium RAD Disks



Method Summary: Test Method TC-196

Rapid Determination of Technetium-99 in Water Using Empore™ Technetium RAD Disks*

Product Description

Empore™ Technetium RAD Disks provide an efficient alternative to conventional radiochemical sample preparation methods that use wet chemistry or packed columns. A proprietary process is used to entrap radium selective adsorbent particles into a matrix of inert PTFE to create a mechanically stable sorbent disk. The disks can be used for purification and concentration of radium from aqueous samples.

Empore technetium RAD disks provide a sample prep solution for large volume aqueous samples and the disk format provides a large surface area for sorbent/sample contact.

Scope and Application

This product is used in place of liquid-liquid extraction, packed columns, or other wet chemistry techniques to quantitatively extract technetium from water samples.

Method Overview

A water sample is passed through an Empore technetium RAD disk with no prior preparation of the sample or the disk. The loaded disk may be dried and placed into a planchet for direct counting in a gas proportional counter or placed wet in a scintillation vial and counted with a liquid scintillation counter.

Interferences

Most ions found in drinking water, surface water and ground water do not interfere with technetium adsorption. Nitrate levels up to 2000 mg/L do not interfere when disks are counted by liquid scintillation.

Safety

No significant safety concerns are associated with this method other than the normal precautions needed for handling radioactive materials.

Apparatus and Materials

- 47 mm Empore technetium RAD disks
- 47 mm single or multiple station vacuum manifold with filter support
- Beta detector – (low background gas flow proportional counter or liquid scintillation counter)
- 50 mm planchets or liquid scintillation vials with scintillation cocktail

Reagents

No reagents are required for this procedure.

Sample Collection, Preservation, and Handling

No special collection, preservation or handling steps are necessary for this procedure. No “wetting” of the disk with methanol or other alcohol is necessary. Samples should not be acidified with nitric acid because of possible interference by nitrate ions.

*This method was developed by Argonne National Laboratories in cooperation with CDS Analytical.

Procedure

1. Measure the sample aliquot. Depending on the activity present and the analytical detection limit required, this may vary from 100 mL or less to several liters.
2. If there is particulate matter present in the sample, use a prefilter to remove any potential beta emitting interfering solids.
3. Mount an Empore technetium RAD disk in the filter support.
4. Pull the sample aliquot through the disk at a nominal flow rate of 50 mL/min. No adjustment of pH or wetting of the disk is necessary.

Counting Options

Disks may be counted by either proportional or scintillation counting.

For use with a scintillation counter, place disk into a vial containing scintillation cocktail.

For use with a gas proportional counter, rinse the disk with 5 mL of acetone while it is still in the manifold, remove the disk from the manifold and dry in a 60°C oven for 15 minutes.

Reference

S. C. Goheen, editor, *DOE Methods for Evaluating Environmental and Waste Management Samples*.

Battelle Press, Columbus, OH (1997), Method RS551, Rapid isolation and measurement of technetium-99 using CDS Analytical Empore™ Technetium Rad Disks.

Important Notice

CDS Analytical MAKES NO WARRANTIES, EXPRESS OR IMPLIED, INCLUDING, BUT NOT LIMITED TO, ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. Since a variety of factors can affect the use and performance of a CDS Analytical product in a particular application, some of which are uniquely within the user's knowledge and control, user is responsible for determining whether or not the CDS Analytical product is fit for a particular purpose and suitable for user's method of application.

Warranty

Seller warrants its equipment against defects in workmanship and material for a period of 12 months from date of shipment from the factory under normal use and service and otherwise when such equipment is used in accordance with instructions furnished by Seller and for purposes disclosed in writing at the time of purchase, if any. Any unauthorized alteration or modification of the equipment by Buyer will void this warranty. Seller's liability under this warranty shall be limited to the replacement or repair, F.O.B., point of manufacture, of any defective equipment or part which, having been returned to the factory, transportation charges prepaid, has been inspected and determined by Seller to be defective. THIS WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EITHER EXPRESSED OR IMPLIED, AS TO DESCRIPTION, QUALITY, MERCHANTABILITY, FITNESS FOR ANY PARTICULAR PURPOSE OR USE, OR ANY OTHER MATTER. Under no circumstances shall Seller be liable to Buyer or any third party for any loss of profits or other direct or indirect costs, expenses, losses or consequential damages arising out of or as a result of any defects in or failure of its products or any part or parts thereof arising out of or as a result of parts or components incorporated in Seller's equipment but not supplied by the Seller.

Technical Information and Data

Technical information and data, recommendations, and other statements provided by CDS Analytical are based on information, tests, or experience which CDS Analytical believes to be reliable, but the accuracy or completeness of such information is not guaranteed. Such technical information and data are intended for persons with knowledge and technical skills sufficient to assess and apply their own informed judgment to the information. No license under any CDS Analytical or third party intellectual property rights is granted or implied with this information.

Safety and Disposal

Always follow safe handling practices while processing Empore products as applicable to your use. Always wear appropriate personal protective equipment while processing Empore products as applicable to your use. Dispose of used Empore products per requirements of your process. Alteration or addition of other materials to Empore products may require unique disposal protocols. Always dispose of used product in accordance with federal, state and local requirements.

Safety Information

WARNING:	Indicates a hazardous situation which, if not avoided, could result in death or serious injury.
WARNING:	
To reduce the risks associated with environmental contamination, handling radiochemical materials or false negative results:	
<ul style="list-style-type: none"> • Read, understand and follow all recommendations in the product MSDS and Instructions for Use. 	
To reduce the risks associated with radiochemical materials:	
<ul style="list-style-type: none"> • Always wear the proper PPE (Personal Protection Equipment) when handling used product. 	
To reduce the risks associated with environmental contamination:	
<ul style="list-style-type: none"> • Dispose of used disks according to standard radiochemical procedures and regulations. Unused disks may be disposed of in accordance with standard procedure for laboratory waste. 	

Note: Empore Solid Phase Extraction Products are intended for solid phase extraction during scientific research only. These products are not intended for use in medical devices or in assessment and treatment of clinical patients.



465 Limestone Rd, Oxford
Toll Free: 1-800-541-6593

Empore is a trademark of CDS Analytical. Used under license.
© CDS Analytical LLC 2018.
All Rights Reserved.
Printed in the USA.
GN-18 10/23 5k